

CORRIGENDUM-1

**Ref: Tender No.: NCPOR/PS/LAB-4130/GT-03.
SUPPLY, INSTALLATION, TESTING & COMMISSIONING OF AUTOMATIC
WEATHER STATION.**

Sr. Nos. as per Tender Technical Specification(Page No. 69 & 70)	Tender Technical Specifications	Revised Tender Technical Specifications
	Data Logger	
4.	Watchdog Timer: Required (auto reset and restore after reset/power glitch)	Supports programmable routines
	Net Radiation Sensor	
50.	Flux Range: 10–1400 W/m ² or better	Omitted
51.	Accuracy: 5 W/m ² or better	Sensitivity: 7 to 15 μ V/W/m ² or better (short wave)
52.	Resolution: 0.5 W/m ² or better	Omitted
54.	Flux Range: 100–700 W/m ² or better	Omitted
55.	Accuracy: 10 W/m ² or better	Sensitivity: 5 to 10 μ V/W/m ² or better (long wave)
56.	Resolution: 0.5 W/m ² or better	Omitted